Quality Biological

PRODUCT: WATER, Molecular Biology

CATALOG#: 351-029-131 Size: 1000ml

LOT#: 724109

CERTIFICATE OF ANALYSIS

This product is de-ionized by reverse osmosis and sterile filtered.

	Specifications	Results
RESISTIVITY:	> 15.11 Meg Ohms/cm	pass
APPEARANCE:	Clear, colorless liquid	pass
DNASE:	Not detectable	pass
RNASE:	Not detectable	pass
PROTEASE:	Not detectable	pass
ENDOTOXIN:	<0.06 Eu/ml	pass
MYCOPLASMA:	Negative	pass
STERILITY:	No evidence of bacterial or fungal contamination after 14 days testing on Fluid Thioglycollate and Tryptic Soy Broth.	pass
RECOMMENDED STORAGE CONDITIONS:	15 – 30°C	
MANUFACTURED DATE:	1/2021	
EXPIRATION DATE:	1/2024	

Michael J. Fisher 02/11/2021

Quality Assurance Date

FOR RESEARCH USE ONLY

7581 Lindbergh Drive ● Gaithersburg, MD 20879
(301) 840-9331 ● Toll Free (800) 443-9331 ● Fax (301) 840-0743
Email: customerservice@qualitybiological.com ● Web Site: www.qualitybiological.com

NOT FOR IN-VITRO DIAGNOSTIC USE.

7581 Lindbergh Drive ● Gaithersburg, MD 20879
(301) 840-9331 ● Toll Free (800) 443-9331 ● Fax (301) 840-0743
Email: customerservice@qualitybiological.com ● Web Site: www.qualitybiological.com