Quality Biological
MAKING SOLUTIONS WORK

PRODUCT: WATER, DEPC TREATED

CATALOG#: 351-068-721 Size: 100ml

LOT#: 723504

CERTIFICATE OF ANALYSIS

This product is ultra pure, de-ionized by reverse osmosis and sterile filtered.

	Specifications	Results
RESISTIVITY:	> 15.11 Meg Ohms/cm	pass
APPEARANCE:	Clear, colorless liquid	pass
DNASE:	Not detectable	pass
RNASE:	Not detectable	pass
PROTEASE:	Not detectable	pass
ENDOTOXIN:	<0.06 Eu/ml	pass
MYCOPLASMA:	Negative	pass
STERILITY:	No evidence of bacterial or fungal contamination after 14 days testing on Fluid Thioglycollate and Tryptic Soy Broth.	pass

RECOMMENDED STORAGE CONDITIONS: 15 – 30°C

MANUFACTURED DATE: 10/2019

EXPIRATION DATE: 10/2022

11/14/2019

Quality Assurance Date

FOR RESEARCH USE ONLY
NOT FOR IN-VITRO DIAGNOSTIC USE

7581 Lindbergh Drive ● Gaithersburg, MD 20879
(301) 840-9331 ● Toll Free (800) 443-9331 ● Fax (301) 840-0743
Email: customerservice@qualitybiological.com ● Web Site: www.qualitybiological.com