Quality Biological
MAKING SOLUTIONS WORK

PRODUCT: PBS pH 7.2
Phosphate Buffered Saline

CATALOG#: 114-056-491 Size: 4L

LOT#: 723966

CERTIFICATE OF ANALYSIS

	Range	Result
pH:	7.1 – 7.3	7.165
OSMOLALITY:	285 - 315 mOsm/kg of water	299
ENDOTOXIN:	Test & Report	<0.010 EU/ml
MYCOPLASMA:	Negative	pass
APPEARANCE:	Clear colorless fluid.	pass
STERILITY:	No evidence of bacterial or fungal contamination after 14 days testing on fluid thioglycollate and Tryptic Soy broth.	pass

RECOMMENDED STORAGE CONDITIONS: 15 – 30°C

MANUFACTURED DATE: 9/2020

EXPIRATION DATE: 9/2022

Ta'Shemma Jackson 10/21/2020

Quality Assurance Date

FOR RESEARCH USE ONLY
NOT FOR IN-VITRO DIAGNOSTIC USE.

7581 Lindbergh Drive ● Gaithersburg, MD 20879
(301) 840-9331 ● Toll Free (800) 443-9331 ● Fax (301) 840-0743
Email: customerservice@qualitybiological.com ● Web Site: www.qualitybiological.com