

PRODUCT: HBSS
Hank's Balanced Salt Solution
without PHENOL RED

CATALOG#: 114-061-101 Size: 500ml

LOT#: 723942

CERTIFICATE OF ANALYSIS

	<u>Range</u>	<u>Result</u>
pH:	6.9 – 7.3	7.072
OSMOLALITY:	266 - 294 mOsm/kg of water	282
ENDOTOXIN:	Test & Report	0.089 EU/ml
MYCOPLASMA:	Negative	pass
APPEARANCE:	Clear colorless fluid	pass
STERILITY:	No evidence of bacterial or fungal contamination after 14 days testing on fluid thioglycollate and Tryptic Soy broth.	pass
RECOMMENDED		
STORAGE CONDITIONS:	15 – 30°C	
MANUFACTURED DATE:	9/2020	
EXPIRATION DATE:	9/2022	

Michael J. Fisher

9/22/2020

Quality Assurance

Date

FOR RESEARCH USE ONLY
NOT FOR IN-VITRO DIAGNOSTIC USE.