

**PRODUCT:** WATER, DEPC TREATED  
**CATALOG#:** 351-068-131      Size: 1000ml  
**LOT#:** 723930

---

### **CERTIFICATE OF ANALYSIS**

---

This product is ultra pure, de-ionized by reverse osmosis and sterile filtered.

	<u>Specifications</u>	<u>Results</u>
<b>RESISTIVITY:</b>	> 15.11 Meg Ohms/cm	pass
<b>APPEARANCE:</b>	Clear, colorless liquid	pass
<b>DNASE:</b>	Not detectable	pass
<b>RNASE:</b>	Not detectable	pass
<b>PROTEASE:</b>	Not detectable	pass
<b>ENDOTOXIN:</b>	<0.06 Eu/ml	pass
<b>MYCOPLASMA:</b>	Negative	pass
<b>STERILITY:</b>	No evidence of bacterial or fungal contamination after 14 days testing on Fluid Thioglycollate and Tryptic Soy Broth.	pass
<b>RECOMMENDED STORAGE CONDITIONS:</b>	15 – 30°C	
<b>MANUFACTURED DATE:</b>	8/2020	
<b>EXPIRATION DATE:</b>	8/2023	

*Michael J. Fisher*

9/17/2020

---

Quality Assurance

Date

**FOR RESEARCH USE ONLY**  
**NOT FOR IN-VITRO DIAGNOSTIC USE**