

**PRODUCT:** WATER, PCR Qualified  
DEPC Treated  
**CATALOG#:** 351-161-671      Size: 10x10ml  
**LOT#:** 723883

---

### **CERTIFICATE OF ANALYSIS**

---

This product is ultra pure, de-ionized by reverse osmosis and sterile filtered.

	<u>Specifications</u>	<u>Results</u>
<b>RESISTIVITY:</b>	> 15.11 Meg Ohms/cm	pass
<b>APPEARANCE:</b>	Clear, colorless liquid	pass
<b>DNASE:</b>	Not detectable	pass
<b>RNASE:</b>	Not detectable	pass
<b>PROTEASE:</b>	Not detectable	pass
<b>ENDOTOXIN:</b>	<0.06 Eu/ml	pass
<b>MYCOPLASMA:</b>	Negative	pass
<b>PCR:</b>	Not detectable for human, mouse and rat DNA contamination.	pass
<b>STERILITY:</b>	No evidence of bacterial or fungal contamination after 14 days testing on fluid thioglycollate and Tryptic Soy broth.	pass
<b>RECOMMENDED STORAGE CONDITIONS:</b>	15 – 30°C	
<b>MANUFACTURED DATE:</b>	7/2020	
<b>EXPIRATION DATE:</b>	7/2023	

*Chanika Johnson*

8/17/2020

---

Quality Assurance

Date

**FOR RESEARCH USE ONLY**

**NOT FOR IN-VITRO DIAGNOSTIC USE.**