

PRODUCT: WATER, Molecular Biology
CATALOG#: 351-029-491 **Size:** 4L
LOT#: 723114

CERTIFICATE OF ANALYSIS

This product is de-ionized by reverse osmosis and sterile filtered.


	<u>Specifications</u>	<u>Results</u>
RESISTIVITY:	> 15.11 Meg Ohms/cm	pass
APPEARANCE:	Clear, colorless liquid	pass
DNASE:	Not detectable	pass
RNASE:	Not detectable	pass
PROTEASE:	Not detectable	pass
ENDOTOXIN:	<0.06 Eu/ml	pass
MYCOPLASMA:	Negative	pass
STERILITY:	No evidence of bacterial or fungal contamination after 14 days testing on Fluid Thioglycollate and Tryptic Soy Broth.	pass

RECOMMENDED

STORAGE CONDITIONS: 15 – 30°C

MANUFACTURED DATE: 2/2019

EXPIRATION DATE: 2/2022


Quality Assurance

3/8/2019

Date

**FOR RESEARCH USE ONLY
NOT FOR IN-VITRO DIAGNOSTIC USE.**