Quality Biological
MAKING SOLUTIONS WORK

PRODUCT: DPBS 1X
Dulbecco's Phosphate Buffer Saline
with Calcium & Magnesium

CATALOG#: 114-059-101 Size: 500ml

LOT#: 723180

CERTIFICATE OF ANALYSIS

	Range	Result
pH:	7.2 – 7.4	7.226
OSMOLALITY :	266 - 294 mOsm/Kg H_2O	289 mOsm/Kg H_2O
ENDOTOXIN:	Test & Report	<0.01 EU/ml
MYCOPLASMA:	Negative	pass
APPEARANCE:	Clear colorless fluid.	pass
STERILITY:	No evidence of bacterial or fungal contamination after 14 days testing on fluid thioglycollate and Tryptic Soy broth.	pass

RECOMMENDED STORAGE CONDITIONS: 2 – 8°C

MANUFACTURED DATE: 3/2019

EXPIRATION DATE: 3/2021

Quality Assurance 4/17/2019
Date

FOR RESEARCH USE ONLY
NOT FOR IN-VITRO DIAGNOSTIC USE.

7581 Lindbergh Drive ● Gaithersburg, MD 20879
(301) 840-9331 ● Toll Free (800) 443-9331 ● Fax (301) 840-0743
Email: customerservice@qualitybiological.com ● Web Site: www.qualitybiological.com