

PRODUCT: Sephadex G-100 Superfine column
50% slurry in Water/0.05% Sodium Azide

CATALOG#: 356-368-466 **Size:** 4ml

LOT#: 723301

CERTIFICATE OF ANALYSIS

	Specification	Results
Measurement of Wet Resin Bed Height	45 ± 2.5	$43.78 \text{ mm} \pm 0.16$


Value reported is MEAN $\pm$ SEM (n=30, 10 beginning, 10 middle, 10 end of the run)

Number	Result	Number	Result
1	44.46	16	44.22
2	44.24	17	43.02
3	44.10	18	43.28
4	43.57	19	43.63
5	43.88	20	44.83
6	43.81	21	42.88
7	43.31	22	44.59
8	43.31	23	45.21
9	42.54	24	45.05
10	42.75	25	45.49
11	42.60	26	44.99
12	43.16	27	44.66
13	42.86	28	43.45
14	42.51	29	42.58
15	44.74	30	43.58
Average	43.78		
SEM	0.16		

RECOMMENDED

STORAGE CONDITIONS: 2-8°C

MANUFACTURED DATE: 6/14/2019



Quality Assurance

6/24/2019

Date

FOR RESEARCH USE ONLY

NOT FOR IN-VITRO DIAGNOSTIC USE.