PRODUCT: ACK LYSING BUFFER

CATALOG#: 118-156-721 Size: 100ml

LOT#: 723857

CERTIFICATE OF ANALYSIS

	Range	Result
pH:	7.2 – 7.6	7.535
OSMOLALITY :	276 - 305 mOsm/Kg H_2O	305
APPEARANCE:	Clear colorless fluid.	pass
STERILITY:	No evidence of bacterial or fungal contamination after 14 days testing on fluid thioglycollate and Tryptic Soy Broth.	pass

RECOMMENDED STORAGE CONDITIONS: 15 – 30°C

MANUFACTURED DATE: 7/2020

EXPIRATION DATE: 7/2022

Chanika Johnson 8/19/2020

Quality Assurance Date

FOR RESEARCH USE ONLY
NOT FOR IN-VITRO DIAGNOSTIC USE.

7581 Lindbergh Drive ● Gaithersburg, MD 20879
(301) 840-9331 ● Toll Free (800) 443-9331 ● Fax (301) 840-0743
Email: customerservice@qualitybiological.com ● Web Site: www.qualitybiological.com