PRODUCT: HBSS
Hank's Balanced Salt Solution
without Calcium & Magnesium

CATALOG#: 114-052-101 Size: 500ml

LOT#: 723766

CERTIFICATE OF ANALYSIS

	Range	Result
pH:	6.9 – 7.3	6.952
OSMOLALITY:	266 - 294 mOsm/kg of water	280
ENDOTOXIN:	Test & Report	<0.020 EU/ml
MYCOPLASMA:	Negative	pass
APPEARANCE:	Pink to a reddish-orange colored fluid.	pass
STERILITY:	No evidence of bacterial or fungal contamination after 14 days testing on fluid thioglycollate and Tryptic Soy broth.	pass
RECOMMENDED STORAGE CONDITIONS:	15 – 30°C	
MANUFACTURED DATE:	4/2020	
EXPIRATION DATE:	4/2022	

Quality Assurance 5/27/2020
Date

FOR RESEARCH USE ONLY
NOT FOR IN-VITRO DIAGNOSTIC USE.

7581 Lindbergh Drive ● Gaithersburg, MD 20879
(301) 840-9331 ● Toll Free (800) 443-9331 ● Fax (301) 840-0743
Email: customerservice@qualitybiological.com ● Web Site: www.qualitybiological.com