


PRODUCT: LB BROTH
LURIA-BERTANI, Miller

CATALOG #: 340-004-131 Size: 1000ml

LOT #: 723757

CERTIFICATE OF ANALYSIS

	<u>Specifications</u>	<u>Results</u>
pH:	7.0 – 7.4	7.190
OSMOLALITY:	389 - 431 mOsm/Kg of water	410
APPEARANCE:	Light amber, clear to slightly opalescent liquid	pass
STERILITY:	Sterilized using saturated steam under pressure	pass
PERFORMANCE:	Passed E. coli growth performance test.	pass
RECOMMENDED STORAGE CONDITIONS:	15 – 30°C	
MANUFACTURED DATE:	4/2020	
EXPIRATION DATE:	4/2022	


Quality Assurance

4/29/2020

Date

**FOR RESEARCH USE ONLY
NOT FOR IN-VITRO DIAGNOSTIC USE.**